

DBH
mmB

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 7/22/2015

Instrument Serial Number: 010603

As of 22-Jul-15

Start Date	End Date	Agency Name
21-Jul-15		Chesapeake PD
01-Jul-15	21-Jul-15	DFS Central Lab
28-May-15	01-Jul-15	Rappahannock Shenandoah Warren Regional Jail
19-Dec-14	28-May-15	DFS Central Lab
16-Dec-14	19-Dec-14	Chesterfield County PD
26-Aug-14	16-Dec-14	DFS Central Lab
31-Mar-14	26-Aug-14	Fairfax County PD
25-Sep-13	31-Mar-14	DFS Central Lab
14-Mar-13	25-Sep-13	Fairfax County PD
05-Nov-12	14-Mar-13	DFS Central Lab
24-May-12	05-Nov-12	Virginia Beach PD
27-Mar-12	24-May-12	DFS Central Lab
22-Nov-11	27-Mar-12	Richmond PD
13-Sep-10	22-Nov-11	DFS Central Lab
04-Jun-10	13-Sep-10	Intoximeters, Inc.
12-May-09	04-Jun-10	DFS Central Lab
18-Mar-09	12-May-09	Norfolk PD
04-Aug-08	18-Mar-09	DFS Central Lab

DBH

mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010603 Worksheet Start Date 7/21/2015

Location Chesapeake Jail

Address 304 Albermarle Dr Chesapeake VA 23322

DFS Technician Donald Hall License No. 21250

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 752 Reference Barometer (mm HG) 752

Reference Barometer(RB)Serial # 009111 RB Calibration Due 5/18/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.291
0.300	0.296	0.287	0.304	Sample 2	0.291
Precision		sample min	sample max	Sample 3	0.290
0.001		0.290	0.291		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.096
0.100	0.099	0.096	0.102	Sample 2	0.096
Precision		sample min	sample max	Sample 3	0.096
0		0.096	0.096		

Dry gas standard Lot No. (with tank no.) AG509101-37

☐ Replaced dry gas standard (+O-ring)

☒ Installed at Location

☐ Removed to DFS-Central

Supplies

Mouthpieces

Certificates of Analysis

Operator Worksheet

Other:

100

100

Notes:

Instrument Serial Number 010603

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Douglas B. Hall

Date

7/22/15

Issuing Analyst

Michelle Bell

Date

*7/22/15**JPN*

DBH

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

mmB

Serial Number: 010603 Test Number: 1445

Test Date: 07/21/2015 Test Time: 14:54 EDT

Dry Gas Target: 0.296

Lot Number: AG506501-07 Exp Date: 03/06/2017

Tank Pressure: 350 psi Barometric Pressure: 751 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	14:54
CHK	0.300	14:55
BLK	0.000	14:57
CHK	0.292	14:58
BLK	0.000	14:59
CHK	0.292	15:00

Test Status: Success

Calibration CRC: AD4641EA

OUT OF TOLERANCE PER QA WORKSHEET

DBH

7/21/15

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010603 Test Number: 1446

Test Date: 07/21/2015 Test Time: 15:02 EDT

Dry Gas Target: 0.296

Lot Number: AG506501-07 Exp Date: 03/06/2017

Tank Pressure: 340 psi Barometric Pressure: 751 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	15:03
CHK	0.291	15:04
BLK	0.000	15:05
CHK	0.291	15:06
BLK	0.000	15:07
CHK	0.290	15:08

Test Status: Success

Calibration CRC: AD4641EA

DBH

mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010603 Test Number: 1447

Test Date: 07/21/2015 Test Time: 15:12 EDT

Dry Gas Target: 0.099

Lot Number: AG509101-37 Exp Date: 04/01/2017

Tank Pressure: 1036 psi Barometric Pressure: 751 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	15:12
CHK	0.096	15:13
BLK	0.000	15:15
CHK	0.096	15:15
BLK	0.000	15:17
CHK	0.096	15:17

Test Status: *Success*

Calibration CRC: AD4641EA

DBH

mmb



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED	NAME OF COURT
INSTRUMENT, TEST,	DFS

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY		AGENCY
HALL, DONALD, B		DFS Central Lab
DFS LICENSE NUMBER	LICENSE EXPIRES	DATE TEST CONDUCTED
21250	10/01/2016	07/21/2015
TEST EQUIPMENT NUMBER		
010603		

RESULTS: TIME SAMPLE TAKEN 15:28 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE

FOR COURT

DBH



IntoxNet MIS Report

Report Generated 22 Jul 2015 at 09:13

mmB

Test Results

Instrument Serial Number 010603

Test # 001448 Subject Test

Test Location 1 Department of
Test Date 21 Jul 2015

Test Location 2 Forensic Science
Test Time 15:22
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 21 Jul 2015

End Time 15:29

Result Time 15:28

Result Date 21 Jul 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 15:22

Data Type BLK

Sample Value 0.000

Sample Time 15:23

Data Type CHK

Sample Value 0.096

Sample Time 15:24

Data Type BLK

Sample Value 0.000

Sample Time 15:25

Data Type SUBJ

Sample Value 0.000

Sample Time 15:26

Data Type BLK

Sample Value 0.000

Sample Time 15:27

Data Type SUBJ

Sample Value 0.000

Sample Time 15:28

Data Type BLK

Sample Value 0.000

Sample Time 15:29

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG509101-37

Standard Expiration Date 04/01/2017

Tank Pressure 1026

Barometric Pressure 751 mmHg

Blow Sample Number 1 Blow Duration 3.38 sec

Blow Volume 1994 cc End-of-Blow Time 15:26

Blow Sample Number 2 Blow Duration 3.33 sec

Blow Volume 1936 cc End-of-Blow Time 15:28

Tamper Evident Stamp 3ba12014

Test Status Code 0

Test Status Success